



**TESTIDES**

NORTH AMERICA'S LEADING  
PEPTIDE TESTING LABORATORY

REPORT NUMBER  
NL-TIRZ-10-040226E

GENERATED  
4/15/2026

# Endotoxin Testing Report

Bacterial Endotoxin Test (LAL)

SAMPLE INFORMATION		TEST INFORMATION	
Sample ID:	NL-TIRZ-10-040226	Method:	LAL Gel-Clot (sensitivity 0.125 EU/mL)
Customer:	True North Labs	Status:	Complete
Date Collected:	April 10, 2026		
Date Tested:	April 14, 2026		

## TEST RESULTS

RESULT	INTERPRETATION
< 0.125 EU/mL	Pass — endotoxin not detected

## NOTES

Endotoxin was not detected in Tirzepatide at the validated 1/16 dilution, with a result of <0.125 EU/mL.

Testides provides analytical testing services and reports laboratory results for the submitted sample only. Testides does not recommend, promote, endorse, certify, or validate the personal or clinical use of research chemicals, nor does it provide medical advice. Any risk assessment language is presented solely as a calculation based on client-provided assumptions (including route of administration and maximum daily dose) and must not be interpreted as a determination of safety, suitability for human use, or regulatory compliance for any finished drug.

CERTIFIED BY TESTIDES

testides.com | April 14, 2026 | 4/15/2026



Scan QR